



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

LABORATORY REPORT

April 1, 2016

Rob Nieman
ALS Environmental
4388 Glendale Milford Road
Cincinnati, OH 45242

RE: Keystone Landfill/PADOH/PADEP / 1603761

Dear Rob:

Enclosed are the results of the sample submitted to our laboratory on March 18, 2016. For your reference, these analyses have been assigned our service request number P1601430.

All analyses were performed according to our laboratory's NELAP and DoD-ELAP-approved quality assurance program. The test results meet requirements of the current NELAP and DoD-ELAP standards, where applicable, and except as noted in the laboratory case narrative provided. For a specific list of NELAP and DoD-ELAP-accredited analytes, refer to the certifications section at www.alsglobal.com. Results are intended to be considered in their entirety and apply only to the samples analyzed and reported herein.

If you have any questions, please call me at (805) 526-7161.

Respectfully submitted,

ALS | Environmental

By Kate Aguilera at 12:36 pm, Apr 01, 2016

Kate Aguilera
Project Manager



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

Client: ALS Environmental
Project: Keystone Landfill/PADOH/PADEP / 1603761

Service Request No: P1601430

CASE NARRATIVE

The sample was received intact under chain of custody on March 18, 2016 and was stored in accordance with the analytical method requirements. Please refer to the sample acceptance check form for additional information. The results reported herein are applicable only to the condition of the sample at the time of sample receipt.

Sulfur Analysis

The sample was analyzed for twenty sulfur compounds per ASTM D 5504-12 using a gas chromatograph equipped with a sulfur chemiluminescence detector (SCD). All compounds with the exception of hydrogen sulfide and carbonyl sulfide are quantitated against the initial calibration curve for methyl mercaptan. This method is included on the laboratory's NELAP scope of accreditation, however it is not part of the DoD-ELAP or AIHA-LAP accreditation.

Volatile Organic Compound Analysis

The sample was also analyzed for volatile organic compounds in accordance with EPA Method TO-15 from the Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Edition (EPA/625/R-96/010b), January, 1999. This procedure is described in laboratory SOP VOA-TO15. The analytical system was comprised of a gas chromatograph / mass spectrometer (GC/MS) interfaced to a whole-air preconcentrator. This method is included on the laboratory's NELAP and DoD-ELAP scope of accreditation, however it is not part of the AIHA-LAP accreditation. Any analytes flagged with an X are not included on the NELAP or DoD-ELAP accreditation.

The spike recovery of 1,4-Dioxane in the Laboratory Control Sample (LCS) was outside the laboratory generated control criteria. The recovery error equates to a potential high bias. However, the recovery in question was within the method criteria, therefore, the data quality has not been significantly affected. No corrective action was taken.

The container was cleaned, prior to sampling, down to the method reporting limit (MRL) reported for this project. Please note, projects which require reporting below the MRL could have results between the MRL and method detection limit (MDL) that are biased high.

The results of analyses are given in the attached laboratory report. All results are intended to be considered in their entirety, and ALS Environmental (ALS) is not responsible for utilization of less than the complete report.

Use of ALS Environmental (ALS)'s Name. Client shall not use ALS's name or trademark in any marketing or reporting materials, press releases or in any other manner ("Materials") whatsoever and shall not attribute to ALS any test result, tolerance or specification derived from ALS's data ("Attribution") without ALS's prior written consent, which may be withheld by ALS for any reason in its sole discretion. To request ALS's consent, Client shall provide copies of the proposed Materials or Attribution and describe in writing Client's proposed use of such Materials or Attribution. If ALS has not provided written approval of the Materials or Attribution within ten (10) days of receipt from Client, Client's request to use ALS's name or trademark in any Materials or Attribution shall be deemed denied. ALS may, in its discretion, reasonably charge Client for its time in reviewing Materials or Attribution requests. Client acknowledges and agrees that the unauthorized use of ALS's name or trademark may cause ALS to incur irreparable harm for which the recovery of money damages will be inadequate. Accordingly, Client acknowledges and agrees that a violation shall justify preliminary injunctive relief. For questions contact the laboratory.



2655 Park Center Dr., Suite A
 Simi Valley, CA 93065
 T: +1 805 526 7161
 F: +1 805 526 7270
www.alsglobal.com

ALS Environmental – Simi Valley

CERTIFICATIONS, ACCREDITATIONS, AND REGISTRATIONS

Agency	Web Site	Number
AIHA	http://www.aihaaccreditedlabs.org	101661
Arizona DHS	http://www.azdhs.gov/lab/license/env.htm	AZ0694
DoD ELAP	http://www.pjlabs.com/search-accredited-labs	L15-398
Florida DOH (NELAP)	http://www.doh.state.fl.us/lab/EnvLabCert/WaterCert.htm	E871020
Maine DHHS	http://www.maine.gov/dhhs/mecdc/environmental-health/water/dwp-services/labcert/labcert.htm	2014025
Minnesota DOH (NELAP)	http://www.health.state.mn.us/accreditation	977273
New Jersey DEP (NELAP)	http://www.nj.gov/dep/oqa/	CA009
New York DOH (NELAP)	http://www.wadsworth.org/labcert/elap/elap.html	11221
Oregon PHD (NELAP)	http://public.health.oregon.gov/LaboratoryServices/EnvironmentalLaboratoryAccreditation/Pages/index.aspx	4068-003
Pennsylvania DEP	http://www.depweb.state.pa.us/labs	68-03307 (Registration)
Texas CEQ (NELAP)	http://www.tceq.texas.gov/field/qa/env_lab_accreditation.html	T104704413-15-6
Utah DOH (NELAP)	http://www.health.utah.gov/lab/labimp/certification/index.html	CA01627201 5-5
Washington DOE	http://www.ecy.wa.gov/programs/eap/labs/lab-accreditation.html	C946

Analyses were performed according to our laboratory's NELAP and DoD-ELAP approved quality assurance program. A complete listing of specific NELAP and DoD-ELAP certified analytes can be found in the certifications section at www.alsglobal.com, or at the accreditation body's website.

Each of the certifications listed above have an explicit Scope of Accreditation that applies to specific matrices/methods/analytes; therefore, please contact the laboratory for information corresponding to a particular certification.

ALS ENVIRONMENTAL

DETAIL SUMMARY REPORT

Client: ALS Environmental
 Project ID: Keystone Landfill/PADOH/PADEP / 1603761

Service Request: P1601430

Date Received: 3/18/2016
 Time Received: 09:45

ASTM D 5504-12 - Sulfur Can	TO-15 - VOC Cans
-----------------------------	------------------

Client Sample ID	Lab Code	Matrix	Date Collected	Time Collected	Container ID	Pi1 (psig)	Pf1 (psig)	ASTM D 5504-12 - Sulfur Can	TO-15 - VOC Cans
1603761-06A (MVH031716-Summa)	P1601430-001	Air	3/17/2016	09:48	AS01078	-4.53	3.89	X	X

CHAIN-OF-CUSTODY RECORD

Date: 23-Mar-16
 COC ID: 3811
 Due Date: 05-Apr-16

Page 1 of 1

P1601430

Subcontractor:
 ALS Simi Valley
 2655 Park Center Drive
 Suite A
 Simi Valley, CA

TEL: (805) 526-7161
 FAX: (805) 526-7270
 Acct #:

Salesperson

Customer Information		Project Information		Parameter/Method Request for Analysis												
Purchase Order	31-1995	Project Name	1603761	A	Sulfur gases by ASTM 5504 Mod. (A5504)											
Work Order		Project Number		B	TO-15 by GC/MS (ETO-15)											
Company Name	ALS Group USA, Corp	Bill To Company	ALS Group USA, Corp	C												
Send Report To	Rob Nieman	Inv Attn	Accounts Payable	D												
Address	4388 Glendale Milford Rd	Address	4388 Glendale Milford Rd	E												
City/State/Zip	Cincinnati, Ohio 45242-	City/State/Zip	Cincinnati, Ohio 45242-	F												
Phone	(513) 733-5336	Phone	(513) 733-5336	G												
Fax	(513) 733-5347	Fax	(513) 733-5347	H												
eMail Address	alsen.reporting@alsglobal.com	eMail CC		I												
				J												
Sample ID			Matrix	Collection Date	24hr	Bottle	A	B	C	D	E	F	G	H	I	J
1603761-06A (MVH031716-Summa)			Air	22/Mar/2016	(1) SUMMA		X									
Relinquished by:		Date/Time	Received by:		Date/Time	Cooler IDs		Report/QC Level								
Relinquished by:		Date/Time	Received by: AI David		Date/Time 03/18/16 @ 09:45			Std								

Comments:

Sub CoC received for P1601430 via email on 3/23/16-Kate

SUMMA

Air - Chain of Custody Record & Analytical Service Request

Page 1 of 1



2655 Park Center Drive, Suite A
 Simi Valley, California 93065
 Phone (805) 526-7181
 Fax (805) 526-7270

Company Name & Address (Reporting Information)
 Pennsylvania Dept. of Environmental Protection
 2 Public Square
 Wilkes-Barre, PA 18701-1915

Project Manager Roger Bellas
 Phone (570) 826-2511
 Fax (570) 826-5448
 Email Address for Result Reporting RBELLAS@PA.GOV

Requested Turnaround Time in Business Days (Surcharges) please circle
 1 Day (100%) 2 Day (75%) 3 Day (50%) 4 Day (35%) 5 Day (25%) 10-Day-Standard

ALS Project No. 71601430

ALS Contact:

Analysis Method

Project Name Keystone Landfill/PADOH/PADEP
 Project Number
 P.O. # Billing Information
 Sampler (Print & Sign)

Client Sample ID	Laboratory ID Number	Date Collected	Time Collected	Canister ID (Bar code # - AC, SC, etc.)	Flow Controller ID (Bar code # - FC #)	Canister Start Pressure "Hg	Canister End Pressure "Hg/psig	Sample Volume	Legal Seal #	ALS Contact	Analysis Method	Project Requirements (MRLs, OAPP)	
												Chains of Custody Seal: (Circle)	Temperature °C
MVHDS1716	①	3/17/16	0948	19034	SFC00072	-29.5	-10		106553			INTACT	3/17/16
												BROKEN	0948
												ABSENT	3/18/16

Report Tier Levels - please select
 Tier I - Results (Default in not specified) _____
 Tier II (Results + QC Summaries) _____
 Tier III (Results + QC & Calibration Summaries) _____
 Tier IV (Date Validation Package) 10% Surcharge _____

Relinquished by: (Signature) Brandon Thour #1727
 Date: 3/17/16 Time: 1150
 Relinquished by: (Signature) [Signature]
 Date: 3/18/16 Time: 0948

Received by: (Signature) [Signature]
 Date: 3/18/16 Time: 0948
 Received by: (Signature) [Signature]
 Date: 3/18/16 Time: 0948

**ALS Environmental
Sample Acceptance Check Form**

Client: ALS Environmental Work order: P1601430
 Project: Keystone Landfill/PADOH/PADEP / 1603761
 Sample(s) received on: 3/18/16 Date opened: 3/18/16 by: ADAVID

Note: This form is used for all samples received by ALS. The use of this form for custody seals is strictly meant to indicate presence/absence and not as an indication of compliance or nonconformity. Thermal preservation and pH will only be evaluated either at the request of the client and/or as required by the method/SOP.

- | | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 1 Were sample containers properly marked with client sample ID? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 2 Did sample containers arrive in good condition? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3 Were chain-of-custody papers used and filled out? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 4 Did sample container labels and/or tags agree with custody papers? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5 Was sample volume received adequate for analysis? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 6 Are samples within specified holding times? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 7 Was proper temperature (thermal preservation) of cooler at receipt adhered to? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 8 Were custody seals on outside of cooler/Box/Container? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| Location of seal(s)? _____ Sealing Lid? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| Were signature and date included? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Were seals intact? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 9 Do containers have appropriate preservation , according to method/SOP or Client specified information? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Is there a client indication that the submitted samples are pH preserved? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were VOA vials checked for presence/absence of air bubbles? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Does the client/method/SOP require that the analyst check the sample pH and <u>if necessary</u> alter it? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 10 Tubes: Are the tubes capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 11 Badges: Are the badges properly capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Are dual bed badges separated and individually capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

Lab Sample ID	Container Description	Required pH *	Received pH	Adjusted pH	VOA Headspace (Presence/Absence)	Receipt / Preservation Comments
P1601430-001.01	6.0 L Silonite Can					

Explain any discrepancies: (include lab sample ID numbers): _____

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: ALS Environmental
Client Sample ID: 1603761-06A (MVH031716-Summa)
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
 ALS Sample ID: P1601430-001

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 6.0 L Silonite Canister
 Test Notes:
 Container ID: AS01078

Date Collected: 3/17/16
 Time Collected: 09:48
 Date Received: 3/18/16
 Date Analyzed: 3/22/16
 Time Analyzed: 15:51
 Volume(s) Analyzed: 1.0 ml(s)

Initial Pressure (psig): -4.53 Final Pressure (psig): 3.89

Canister Dilution Factor: 1.83

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	13	ND	9.2	
463-58-1	Carbonyl Sulfide	ND	22	ND	9.2	
74-93-1	Methyl Mercaptan	ND	18	ND	9.2	
75-08-1	Ethyl Mercaptan	ND	23	ND	9.2	
75-18-3	Dimethyl Sulfide	ND	23	ND	9.2	
75-15-0	Carbon Disulfide	ND	14	ND	4.6	
75-33-2	Isopropyl Mercaptan	ND	28	ND	9.2	
75-66-1	tert-Butyl Mercaptan	ND	34	ND	9.2	
107-03-9	n-Propyl Mercaptan	ND	28	ND	9.2	
624-89-5	Ethyl Methyl Sulfide	ND	28	ND	9.2	
110-02-1	Thiophene	ND	31	ND	9.2	
513-44-0	Isobutyl Mercaptan	ND	34	ND	9.2	
352-93-2	Diethyl Sulfide	ND	34	ND	9.2	
109-79-5	n-Butyl Mercaptan	ND	34	ND	9.2	
624-92-0	Dimethyl Disulfide	ND	18	ND	4.6	
616-44-4	3-Methylthiophene	ND	37	ND	9.2	
110-01-0	Tetrahydrothiophene	ND	33	ND	9.2	
638-02-8	2,5-Dimethylthiophene	ND	42	ND	9.2	
872-55-9	2-Ethylthiophene	ND	42	ND	9.2	
110-81-6	Diethyl Disulfide	ND	23	ND	4.6	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 1

Client: ALS Environmental
Client Sample ID: Method Blank
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
 ALS Sample ID: P160322-MB

Test Code: ASTM D 5504-12
 Instrument ID: Agilent 7890A/GC22/SCD
 Analyst: Mike Conejo
 Sample Type: 6.0 L Silonite Canister
 Test Notes:

Date Collected: NA
 Time Collected: NA
 Date Received: NA
 Date Analyzed: 3/22/16
 Time Analyzed: 07:47
 Volume(s) Analyzed: 1.0 ml(s)

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
7783-06-4	Hydrogen Sulfide	ND	7.0	ND	5.0	
463-58-1	Carbonyl Sulfide	ND	12	ND	5.0	
74-93-1	Methyl Mercaptan	ND	9.8	ND	5.0	
75-08-1	Ethyl Mercaptan	ND	13	ND	5.0	
75-18-3	Dimethyl Sulfide	ND	13	ND	5.0	
75-15-0	Carbon Disulfide	ND	7.8	ND	2.5	
75-33-2	Isopropyl Mercaptan	ND	16	ND	5.0	
75-66-1	tert-Butyl Mercaptan	ND	18	ND	5.0	
107-03-9	n-Propyl Mercaptan	ND	16	ND	5.0	
624-89-5	Ethyl Methyl Sulfide	ND	16	ND	5.0	
110-02-1	Thiophene	ND	17	ND	5.0	
513-44-0	Isobutyl Mercaptan	ND	18	ND	5.0	
352-93-2	Diethyl Sulfide	ND	18	ND	5.0	
109-79-5	n-Butyl Mercaptan	ND	18	ND	5.0	
624-92-0	Dimethyl Disulfide	ND	9.6	ND	2.5	
616-44-4	3-Methylthiophene	ND	20	ND	5.0	
110-01-0	Tetrahydrothiophene	ND	18	ND	5.0	
638-02-8	2,5-Dimethylthiophene	ND	23	ND	5.0	
872-55-9	2-Ethylthiophene	ND	23	ND	5.0	
110-81-6	Diethyl Disulfide	ND	12	ND	2.5	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 1

Client: ALS Environmental
Client Sample ID: Lab Control Sample
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
ALS Sample ID: P160322-LCS

Test Code: ASTM D 5504-12
Instrument ID: Agilent 7890A/GC22/SCD
Analyst: Mike Conejo
Sample Type: 6.0 L Silonite Canister
Test Notes:

Date Collected: NA
Date Received: NA
Date Analyzed: 3/22/16
Volume(s) Analyzed: NA ml(s)

CAS #	Compound	Spike Amount ppbV	Result ppbV	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
7783-06-4	Hydrogen Sulfide	1,000	893	89	65-138	
463-58-1	Carbonyl Sulfide	1,000	934	93	60-135	
74-93-1	Methyl Mercaptan	1,000	890	89	57-140	

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 3

Client: ALS Environmental
Client Sample ID: 1603761-06A (MVH031716-Summa)
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
 ALS Sample ID: P1601430-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9
 Analyst: Simon Cao
 Sample Type: 6.0 L Silonite Canister
 Test Notes:
 Container ID: AS01078

Date Collected: 3/17/16
 Date Received: 3/18/16
 Date Analyzed: 3/24/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -4.53 Final Pressure (psig): 3.89

Canister Dilution Factor: 1.83

CAS #	Compound	Result	MRL	Result	MRL	Data Qualifier
		µg/m ³	µg/m ³	ppbV	ppbV	
115-07-1	Propene	2.4	0.92	1.4	0.53	
75-71-8	Dichlorodifluoromethane (CFC 12)	2.2	0.92	0.44	0.19	
74-87-3	Chloromethane	ND	0.92	ND	0.44	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	ND	0.92	ND	0.13	
75-01-4	Vinyl Chloride	ND	0.92	ND	0.36	
106-99-0	1,3-Butadiene	ND	0.92	ND	0.41	
74-83-9	Bromomethane	ND	0.92	ND	0.24	
75-00-3	Chloroethane	ND	0.92	ND	0.35	
64-17-5	Ethanol	ND	9.2	ND	4.9	
75-05-8	Acetonitrile	ND	0.92	ND	0.55	
107-02-8	Acrolein	ND	3.7	ND	1.6	
67-64-1	Acetone	ND	9.2	ND	3.9	
75-69-4	Trichlorofluoromethane	1.1	0.92	0.20	0.16	
67-63-0	2-Propanol (Isopropyl Alcohol)	ND	9.2	ND	3.7	
107-13-1	Acrylonitrile	ND	0.92	ND	0.42	
75-35-4	1,1-Dichloroethene	ND	0.92	ND	0.23	
75-09-2	Methylene Chloride	ND	0.92	ND	0.26	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	ND	0.92	ND	0.29	
76-13-1	Trichlorotrifluoroethane	ND	0.92	ND	0.12	
75-15-0	Carbon Disulfide	ND	9.2	ND	2.9	
156-60-5	trans-1,2-Dichloroethene	ND	0.92	ND	0.23	
75-34-3	1,1-Dichloroethane	ND	0.92	ND	0.23	
1634-04-4	Methyl tert-Butyl Ether	ND	0.92	ND	0.25	
108-05-4	Vinyl Acetate	ND	9.2	ND	2.6	
78-93-3	2-Butanone (MEK)	ND	9.2	ND	3.1	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 2 of 3

Client: ALS Environmental
Client Sample ID: 1603761-06A (MVH031716-Summa)
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
 ALS Sample ID: P1601430-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9
 Analyst: Simon Cao
 Sample Type: 6.0 L Silonite Canister
 Test Notes:
 Container ID: AS01078

Date Collected: 3/17/16
 Date Received: 3/18/16
 Date Analyzed: 3/24/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -4.53 Final Pressure (psig): 3.89

Canister Dilution Factor: 1.83

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
156-59-2	cis-1,2-Dichloroethene	ND	0.92	ND	0.23	
141-78-6	Ethyl Acetate	ND	1.8	ND	0.51	
110-54-3	n-Hexane	ND	0.92	ND	0.26	
67-66-3	Chloroform	ND	0.92	ND	0.19	
109-99-9	Tetrahydrofuran (THF)	ND	0.92	ND	0.31	
107-06-2	1,2-Dichloroethane	ND	0.92	ND	0.23	
71-55-6	1,1,1-Trichloroethane	ND	0.92	ND	0.17	
71-43-2	Benzene	ND	0.92	ND	0.29	
56-23-5	Carbon Tetrachloride	ND	0.92	ND	0.15	
110-82-7	Cyclohexane	ND	1.8	ND	0.53	
78-87-5	1,2-Dichloropropane	ND	0.92	ND	0.20	
75-27-4	Bromodichloromethane	ND	0.92	ND	0.14	
79-01-6	Trichloroethene	ND	0.92	ND	0.17	
123-91-1	1,4-Dioxane	ND	0.92	ND	0.25	
80-62-6	Methyl Methacrylate	ND	1.8	ND	0.45	
142-82-5	n-Heptane	ND	0.92	ND	0.22	
10061-01-5	cis-1,3-Dichloropropene	ND	0.92	ND	0.20	
108-10-1	4-Methyl-2-pentanone	ND	0.92	ND	0.22	
10061-02-6	trans-1,3-Dichloropropene	ND	0.92	ND	0.20	
79-00-5	1,1,2-Trichloroethane	ND	0.92	ND	0.17	
108-88-3	Toluene	1.2	0.92	0.31	0.24	
591-78-6	2-Hexanone	ND	0.92	ND	0.22	
124-48-1	Dibromochloromethane	ND	0.92	ND	0.11	
106-93-4	1,2-Dibromoethane	ND	0.92	ND	0.12	
123-86-4	n-Butyl Acetate	ND	0.92	ND	0.19	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 3 of 3

Client: ALS Environmental
Client Sample ID: 1603761-06A (MVH031716-Summa)
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430
 ALS Sample ID: P1601430-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9
 Analyst: Simon Cao
 Sample Type: 6.0 L Silonite Canister
 Test Notes:
 Container ID: AS01078

Date Collected: 3/17/16
 Date Received: 3/18/16
 Date Analyzed: 3/24/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -4.53 Final Pressure (psig): 3.89

Canister Dilution Factor: 1.83

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
111-65-9	n-Octane	ND	0.92	ND	0.20	
127-18-4	Tetrachloroethene	ND	0.92	ND	0.13	
108-90-7	Chlorobenzene	ND	0.92	ND	0.20	
100-41-4	Ethylbenzene	ND	0.92	ND	0.21	
179601-23-1	m,p-Xylenes	ND	1.8	ND	0.42	
75-25-2	Bromoform	ND	0.92	ND	0.089	
100-42-5	Styrene	ND	0.92	ND	0.21	
95-47-6	o-Xylene	ND	0.92	ND	0.21	
111-84-2	n-Nonane	ND	0.92	ND	0.17	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.92	ND	0.13	
98-82-8	Cumene	ND	0.92	ND	0.19	
80-56-8	alpha-Pinene	ND	0.92	ND	0.16	
103-65-1	n-Propylbenzene	ND	0.92	ND	0.19	
622-96-8	4-Ethyltoluene	ND	0.92	ND	0.19	
108-67-8	1,3,5-Trimethylbenzene	ND	0.92	ND	0.19	
95-63-6	1,2,4-Trimethylbenzene	ND	0.92	ND	0.19	
100-44-7	Benzyl Chloride	ND	0.92	ND	0.18	
541-73-1	1,3-Dichlorobenzene	ND	0.92	ND	0.15	
106-46-7	1,4-Dichlorobenzene	ND	0.92	ND	0.15	
95-50-1	1,2-Dichlorobenzene	ND	0.92	ND	0.15	
5989-27-5	d-Limonene	ND	0.92	ND	0.16	
96-12-8	1,2-Dibromo-3-chloropropane	ND	0.92	ND	0.095	
120-82-1	1,2,4-Trichlorobenzene	ND	0.92	ND	0.12	
91-20-3	Naphthalene	ND	0.92	ND	0.17	
87-68-3	Hexachlorobutadiene	ND	0.92	ND	0.086	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 3

Client: ALS Environmental

Client Sample ID: Method Blank

Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

ALS Sample ID: P160324-MB

Test Code: EPA TO-15

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Analyst: Simon Cao

Sample Type: 6.0 L Silonite Canister

Test Notes:

Date Collected: NA

Date Received: NA

Date Analyzed: 3/24/16

Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

CAS #	Compound	Result	MRL	Result	MRL	Data Qualifier
		µg/m ³	µg/m ³	ppbV	ppbV	
115-07-1	Propene	ND	0.50	ND	0.29	
75-71-8	Dichlorodifluoromethane (CFC 12)	ND	0.50	ND	0.10	
74-87-3	Chloromethane	ND	0.50	ND	0.24	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	ND	0.50	ND	0.072	
75-01-4	Vinyl Chloride	ND	0.50	ND	0.20	
106-99-0	1,3-Butadiene	ND	0.50	ND	0.23	
74-83-9	Bromomethane	ND	0.50	ND	0.13	
75-00-3	Chloroethane	ND	0.50	ND	0.19	
64-17-5	Ethanol	ND	5.0	ND	2.7	
75-05-8	Acetonitrile	ND	0.50	ND	0.30	
107-02-8	Acrolein	ND	2.0	ND	0.87	
67-64-1	Acetone	ND	5.0	ND	2.1	
75-69-4	Trichlorofluoromethane	ND	0.50	ND	0.089	
67-63-0	2-Propanol (Isopropyl Alcohol)	ND	5.0	ND	2.0	
107-13-1	Acrylonitrile	ND	0.50	ND	0.23	
75-35-4	1,1-Dichloroethene	ND	0.50	ND	0.13	
75-09-2	Methylene Chloride	ND	0.50	ND	0.14	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	ND	0.50	ND	0.16	
76-13-1	Trichlorotrifluoroethane	ND	0.50	ND	0.065	
75-15-0	Carbon Disulfide	ND	5.0	ND	1.6	
156-60-5	trans-1,2-Dichloroethene	ND	0.50	ND	0.13	
75-34-3	1,1-Dichloroethane	ND	0.50	ND	0.12	
1634-04-4	Methyl tert-Butyl Ether	ND	0.50	ND	0.14	
108-05-4	Vinyl Acetate	ND	5.0	ND	1.4	
78-93-3	2-Butanone (MEK)	ND	5.0	ND	1.7	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 2 of 3

Client: ALS Environmental

Client Sample ID: Method Blank

Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

ALS Sample ID: P160324-MB

Test Code: EPA TO-15

Date Collected: NA

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: NA

Analyst: Simon Cao

Date Analyzed: 3/24/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 1.00 Liter(s)

Test Notes:

Canister Dilution Factor: 1.00

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
156-59-2	cis-1,2-Dichloroethene	ND	0.50	ND	0.13	
141-78-6	Ethyl Acetate	ND	1.0	ND	0.28	
110-54-3	n-Hexane	ND	0.50	ND	0.14	
67-66-3	Chloroform	ND	0.50	ND	0.10	
109-99-9	Tetrahydrofuran (THF)	ND	0.50	ND	0.17	
107-06-2	1,2-Dichloroethane	ND	0.50	ND	0.12	
71-55-6	1,1,1-Trichloroethane	ND	0.50	ND	0.092	
71-43-2	Benzene	ND	0.50	ND	0.16	
56-23-5	Carbon Tetrachloride	ND	0.50	ND	0.080	
110-82-7	Cyclohexane	ND	1.0	ND	0.29	
78-87-5	1,2-Dichloropropane	ND	0.50	ND	0.11	
75-27-4	Bromodichloromethane	ND	0.50	ND	0.075	
79-01-6	Trichloroethene	ND	0.50	ND	0.093	
123-91-1	1,4-Dioxane	ND	0.50	ND	0.14	
80-62-6	Methyl Methacrylate	ND	1.0	ND	0.24	
142-82-5	n-Heptane	ND	0.50	ND	0.12	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	ND	0.11	
108-10-1	4-Methyl-2-pentanone	ND	0.50	ND	0.12	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	ND	0.11	
79-00-5	1,1,2-Trichloroethane	ND	0.50	ND	0.092	
108-88-3	Toluene	ND	0.50	ND	0.13	
591-78-6	2-Hexanone	ND	0.50	ND	0.12	
124-48-1	Dibromochloromethane	ND	0.50	ND	0.059	
106-93-4	1,2-Dibromoethane	ND	0.50	ND	0.065	
123-86-4	n-Butyl Acetate	ND	0.50	ND	0.11	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 3 of 3

Client: ALS Environmental

Client Sample ID: Method Blank

Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

ALS Sample ID: P160324-MB

Test Code: EPA TO-15

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Analyst: Simon Cao

Sample Type: 6.0 L Silonite Canister

Test Notes:

Date Collected: NA

Date Received: NA

Date Analyzed: 3/24/16

Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
111-65-9	n-Octane	ND	0.50	ND	0.11	
127-18-4	Tetrachloroethene	ND	0.50	ND	0.074	
108-90-7	Chlorobenzene	ND	0.50	ND	0.11	
100-41-4	Ethylbenzene	ND	0.50	ND	0.12	
179601-23-1	m,p-Xylenes	ND	1.0	ND	0.23	
75-25-2	Bromoform	ND	0.50	ND	0.048	
100-42-5	Styrene	ND	0.50	ND	0.12	
95-47-6	o-Xylene	ND	0.50	ND	0.12	
111-84-2	n-Nonane	ND	0.50	ND	0.095	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	ND	0.073	
98-82-8	Cumene	ND	0.50	ND	0.10	
80-56-8	alpha-Pinene	ND	0.50	ND	0.090	
103-65-1	n-Propylbenzene	ND	0.50	ND	0.10	
622-96-8	4-Ethyltoluene	ND	0.50	ND	0.10	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	ND	0.10	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	ND	0.10	
100-44-7	Benzyl Chloride	ND	0.50	ND	0.097	
541-73-1	1,3-Dichlorobenzene	ND	0.50	ND	0.083	
106-46-7	1,4-Dichlorobenzene	ND	0.50	ND	0.083	
95-50-1	1,2-Dichlorobenzene	ND	0.50	ND	0.083	
5989-27-5	d-Limonene	ND	0.50	ND	0.090	
96-12-8	1,2-Dibromo-3-chloropropane	ND	0.50	ND	0.052	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	ND	0.067	
91-20-3	Naphthalene	ND	0.50	ND	0.095	
87-68-3	Hexachlorobutadiene	ND	0.50	ND	0.047	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

SURROGATE SPIKE RECOVERY RESULTS

Page 1 of 1

Client: ALS Environmental
Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9
 Analyst: Simon Cao
 Sample Type: 6.0 L Silonite Canister(s)
 Test Notes:

Date(s) Collected: 3/17/16
 Date(s) Received: 3/18/16
 Date(s) Analyzed: 3/24/16

Client Sample ID	ALS Sample ID	1,2-Dichloroethane-d4	Toluene-d8	Bromofluorobenzene	Acceptance Limits	Data Qualifier
		Percent Recovered	Percent Recovered	Percent Recovered		
Method Blank	P160324-MB	98	100	102	70-130	
Lab Control Sample	P160324-LCS	95	98	105	70-130	
1603761-06A (MVH031716-Summa)	P1601430-001	98	98	105	70-130	

Surrogate percent recovery is verified and accepted based on the on-column result.

Reported results are shown in concentration units and as a result of the calculation, may vary slightly from the on-column percent recovery.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 3

Client: ALS Environmental

Client Sample ID: Lab Control Sample

Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

ALS Sample ID: P160324-LCS

Test Code: EPA TO-15

Date Collected: NA

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: NA

Analyst: Simon Cao

Date Analyzed: 3/24/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 0.125 Liter(s)

Test Notes:

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
115-07-1	Propene	196	207	106	49-131	
75-71-8	Dichlorodifluoromethane (CFC 12)	188	184	98	65-117	
74-87-3	Chloromethane	200	193	97	48-132	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	204	182	89	65-122	
75-01-4	Vinyl Chloride	200	187	94	65-128	
106-99-0	1,3-Butadiene	206	216	105	62-143	
74-83-9	Bromomethane	202	201	100	65-130	
75-00-3	Chloroethane	200	206	103	69-126	
64-17-5	Ethanol	998	1090	109	57-126	
75-05-8	Acetonitrile	212	205	97	51-134	
107-02-8	Acrolein	214	240	112	55-146	
67-64-1	Acetone	1,080	1130	105	57-120	
75-69-4	Trichlorofluoromethane	216	191	88	59-139	
67-63-0	2-Propanol (Isopropyl Alcohol)	418	455	109	59-129	
107-13-1	Acrylonitrile	212	240	113	64-136	
75-35-4	1,1-Dichloroethene	216	230	106	72-123	
75-09-2	Methylene Chloride	222	220	99	63-117	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	218	245	112	50-141	
76-13-1	Trichlorotrifluoroethane	220	227	103	68-118	
75-15-0	Carbon Disulfide	210	211	100	55-143	
156-60-5	trans-1,2-Dichloroethene	210	246	117	69-129	
75-34-3	1,1-Dichloroethane	212	220	104	66-122	
1634-04-4	Methyl tert-Butyl Ether	216	229	106	55-128	
108-05-4	Vinyl Acetate	1,040	1270	122	66-140	
78-93-3	2-Butanone (MEK)	220	251	114	62-127	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result. Reported results are shown in concentration units and as a result of the calculation, may vary slightly.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 2 of 3

Client:	ALS Environmental	
Client Sample ID:	Lab Control Sample	ALS Project ID: P1601430
Client Project ID:	Keystone Landfill/PADOH/PADEP / 1603761	ALS Sample ID: P160324-LCS
Test Code:	EPA TO-15	Date Collected: NA
Instrument ID:	Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9	Date Received: NA
Analyst:	Simon Cao	Date Analyzed: 3/24/16
Sample Type:	6.0 L Silonite Canister	Volume(s) Analyzed: 0.125 Liter(s)
Test Notes:		

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
156-59-2	cis-1,2-Dichloroethene	218	239	110	65-125	
141-78-6	Ethyl Acetate	428	512	120	64-132	
110-54-3	n-Hexane	212	231	109	58-126	
67-66-3	Chloroform	224	226	101	68-117	
109-99-9	Tetrahydrofuran (THF)	220	242	110	64-123	
107-06-2	1,2-Dichloroethane	214	223	104	63-124	
71-55-6	1,1,1-Trichloroethane	210	226	108	68-120	
71-43-2	Benzene	226	206	91	61-110	
56-23-5	Carbon Tetrachloride	230	244	106	65-137	
110-82-7	Cyclohexane	424	458	108	68-122	
78-87-5	1,2-Dichloropropane	216	233	108	67-122	
75-27-4	Bromodichloromethane	218	243	111	71-124	
79-01-6	Trichloroethene	216	213	99	71-121	
123-91-1	1,4-Dioxane	210	260	124	67-122	L
80-62-6	Methyl Methacrylate	422	513	122	76-130	
142-82-5	n-Heptane	216	236	109	67-125	
10061-01-5	cis-1,3-Dichloropropene	208	254	122	73-131	
108-10-1	4-Methyl-2-pentanone	220	265	120	66-132	
10061-02-6	trans-1,3-Dichloropropene	210	263	125	76-135	
79-00-5	1,1,2-Trichloroethane	216	243	113	73-121	
108-88-3	Toluene	218	219	100	67-117	
591-78-6	2-Hexanone	220	271	123	59-128	
124-48-1	Dibromochloromethane	220	257	117	73-132	
106-93-4	1,2-Dibromoethane	218	252	116	73-128	
123-86-4	n-Butyl Acetate	226	279	123	61-136	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result. Reported results are shown in concentration units and as a result of the calculation, may vary slightly. L = Laboratory control sample recovery outside the specified limits, results may be biased high.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 3 of 3

Client: ALS Environmental

Client Sample ID: Lab Control Sample

Client Project ID: Keystone Landfill/PADOH/PADEP / 1603761

ALS Project ID: P1601430

ALS Sample ID: P160324-LCS

Test Code: EPA TO-15

Date Collected: NA

Instrument ID: Tekmar AUTOCAN/Agilent 5973inert/6890N/MS9

Date Received: NA

Analyst: Simon Cao

Date Analyzed: 3/24/16

Sample Type: 6.0 L Silonite Canister

Volume(s) Analyzed: 0.125 Liter(s)

Test Notes:

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
111-65-9	n-Octane	210	233	111	67-124	
127-18-4	Tetrachloroethene	202	214	106	65-126	
108-90-7	Chlorobenzene	220	228	104	68-120	
100-41-4	Ethylbenzene	218	236	108	69-123	
179601-23-1	m,p-Xylenes	428	472	110	67-125	
75-25-2	Bromoform	228	251	110	68-153	
100-42-5	Styrene	222	265	119	68-132	
95-47-6	o-Xylene	210	226	108	67-124	
111-84-2	n-Nonane	204	229	112	60-130	
79-34-5	1,1,2,2-Tetrachloroethane	210	253	120	72-128	
98-82-8	Cumene	208	223	107	67-124	
80-56-8	alpha-Pinene	212	239	113	67-129	
103-65-1	n-Propylbenzene	204	224	110	67-125	
622-96-8	4-Ethyltoluene	214	231	108	66-128	
108-67-8	1,3,5-Trimethylbenzene	214	236	110	65-125	
95-63-6	1,2,4-Trimethylbenzene	218	252	116	62-134	
100-44-7	Benzyl Chloride	220	295	134	74-145	
541-73-1	1,3-Dichlorobenzene	228	252	111	63-133	
106-46-7	1,4-Dichlorobenzene	208	238	114	62-129	
95-50-1	1,2-Dichlorobenzene	220	246	112	62-134	
5989-27-5	d-Limonene	210	247	118	66-137	
96-12-8	1,2-Dibromo-3-chloropropane	218	267	122	71-147	
120-82-1	1,2,4-Trichlorobenzene	230	260	113	60-145	
91-20-3	Naphthalene	218	253	116	56-158	
87-68-3	Hexachlorobutadiene	230	247	107	56-139	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result. Reported results are shown in concentration units and as a result of the calculation, may vary slightly.